

Apatite-Forming Ability of Doped-Ceria Coatings for Orthopedic Implants

Authors : Ayda Khosravanihaghighi, Pramod Koshy, Bill Walsh, Vedran Lovric, Charles Christopher Sorrell

Abstract : There is an increasing demand for orthopedic implants owing to the increasing numbers of the aging population. Titanium alloy (Ti6Al4V) is a common material used for orthopedic implants owing to its advantageous properties in terms of good corrosion resistance, minimal elastic modulus mismatch with bone, bio-inertness, and high mechanical strength. However, it is important to improve the bioactivity and osseointegration of the titanium alloy and this can be achieved by coating the implant surface with suitable ceramic materials. In the present work, pure and doped-ceria (CeO₂) coatings were deposited by spin coating on the titanium alloy surface in order to enhance the biological interactions between the surface of the implant and the surrounding tissue. In order to examine the bone-binding ability of an implant, simulated body fluid (SBF) tests were conducted in order to assess the capability of apatite layer formation on the surface and thus predict in vivo bone bioactivity. Characterization was done using scanning electron microscopy (SEM) and X-ray diffraction (XRD) analyses to determine the extent of apatite formation. Preliminary tests showed that the CeO₂ coatings were biocompatible and that the extent of apatite formation and its characteristics can be enhanced by doping with suitable metal ions.

Keywords : apatite layer, biocompatibility, ceria, orthopaedic implant, SBF, spin coater, Ti-implant

Conference Title : ICCB 2019 : International Conference on Ceramic Biomaterials

Conference Location : Tokyo, Japan

Conference Dates : October 07-08, 2019