

Study for a Non-Invasive Method of Respiratory Resistance Measurement among Patients with Airways Obstructions

Authors : Aicha Laouani, Pascale Calabrese, Sonia Rouatbi, Saad Saguem

Abstract : Distances between signals (S d) and between asters (A d) calculated from respiratory inductive plethysmography signals has been used in order to evaluation airways resistances (Raw) during reversibility test among 28 subject with airways obstructions. Correlations studies between these distances and Raw measured by body plethysmography (BP) showed that these RIP variables could be potentially used in airway resistance assessment in patients with airway obstruction. Significant correlation was found between ΔA_d and airway resistance changes (ΔR_{aw}) ($r= 0.407$, $p=0.03$) and not between ΔS_d and ΔR_{aw} . This assumption was supported by the high correlations found when relating the average of ΔS and of ΔA calculated on successive intervals of ΔR_{aw} , with the ΔR_{aw} averages calculated for each interval ($r= 0.892$, $p= 0.006$ and $r= 0.857$, $p=0.006$ respectively).

Keywords : airways obstruction, distances, respiratory inductive plethysmography, reversibility test

Conference Title : ICBM 2014 : International Conference on Biomechanics

Conference Location : Venice, Italy

Conference Dates : June 19-20, 2014