

## **Destruction of Atherosclerotic Plaque Using Pulse Ultrasound with a Planar Rectangular Ultrasound Transducer**

**Authors :** Christakis Damianou, Christos Christofi, Nicos Mylonas

**Abstract :** The aim of the proposed study was to evaluate mechanical mode ultrasound using a flat rectangular (3x10 mm<sup>2</sup>) MRI compatible transducer operating at 5 MHz for destroying atherosclerotic plaque. The system was tested initially in a Hydroxyapatite-polyalactide (HA/PLA) model. An optimized protocol was decided and then applied in atherosclerotic plaque of a rabbit. The plaque in the rabbit was created using a high cholesterol diet. The atherosclerotic plaque was imaged using MRI. This study shows that the destruction of atherosclerotic plaque is feasible.

**Keywords :** mri, ultrasound, atherosclerotic, plaque, pulse

**Conference Title :** ICBE 2014 : International Conference on Biomedical Engineering

**Conference Location :** Bangkok, Thailand

**Conference Dates :** December 18-19, 2014