## Comparison of the Response of TLD-100 and TLD-100H Dosimeters in Diagnostic Radiology

Authors : S. Sina, B. Zeinali, M. Karimipourfard, F. Lotfalizadeh, M. Sadeghi, E. Zamani, M. Zehtabian, R. Faghihi

**Abstract :** Proper dosimetery is very essential in diagnostic radiology. The goal of this study is to verify the application of LiF:Mg, Cu, P (TLD100H) in obtaining the entrance skin dose (ESD) of patients undergoing diagnostic radiology. The results of dosimetry performed by TLD-100H were compared with those obtained by TLD100, which is a common dosimeter in diagnostic radiology. The results show a close agreement between the dose measured by the two dosimeters. According to the results of this study, the TLD-100H dosimeters have higher sensitivities (i.e. signal(nc)/dose) than TLD-100. Therefore, it is suggested that the TLD-100H are effective dosimeters for dosimetry in low dose fields.

Keywords : entrance skin dose, TLD, diagnostic radiology, dosimeter

Conference Title : ICMPRPR 2014 : International Conference on Medical Physics, Radiation Protection and Radiobiology

Conference Location : Istanbul, Türkiye

Conference Dates : September 29-30, 2014