

Determination of Four Anions in the Ground Layer of Tomb Murals by Ion Chromatography

Authors : Liping Qiu, Xiaofeng Zhang

Abstract : The ion chromatography method for the rapid determination of four anions (F^- , Cl^- , SO_4^{2-} , NO_3^-) in burial ground poles was optimized. The $L_9(3^4)$ orthogonal test was used to determine the optimal parameters of sample pretreatment: accurately weigh 2.000g of sample, add 10mL of ultrapure water, and extract for 40min under the conditions of shaking temperature 40°C and shaking speed 180 r·min⁻¹. The eluent was 25 mmol/L KOH solution, the analytical column was Ion Pac® AS11-SH (250 mm × 4.0 mm), and the purified filtrate was measured by a conductivity detector. Under this method, the detection limit of each ion is 0.066~0.078mg/kg, the relative standard deviation is 0.86%~2.44% (n=7), and the recovery rate is 94.6~101.9.

Keywords : ion chromatography, tomb, anion (F^- , Cl^- , SO_4^{2-} , NO_3^-), environmental protection

Conference Title : ICWPCE 2022 : International Conference on Water Pollution Control Engineering

Conference Location : Jerusalem, Israel

Conference Dates : November 29-30, 2022