Design, Construction and Characterization of a 3He Proportional Counter for Detecting Thermal Neutron

Authors : M. Fares, S. Mameri, I. Abdlani, K. Negara

Abstract : Neutron detectors in general, proportional counters gas filling based isotope 3He in particular are going to be essential for monitoring and control of certain nuclear facilities, monitoring of experimentation around neutron beams and channels nuclear research reactors, radiation protection instruments and other tools multifaceted exploration and testing of materials, etc. This work consists of a measurement campaign features two Proportional Counters 3He (3He: LND252/USA CP, CP prototype: 3He LND/DDM). This is to make a comparison study of a CP 3He LND252/USA reference one hand, and in the context of routine periodic monitoring of the characteristics of the detectors for controlling the operation especially for laboratory prototypes. In this paper, we have described the different characteristics of the detectors and the experimental protocols used. Tables of measures have been developed and the different curves were plotted. The experimental campaign at stake: 2 PC 3He were thus characterized: Their characteristics (sensitivity, energy pulse height distribution spectra, gas amplification etc.) Were identified: 01 PC 3He 1" Type: prototype DEDIN/DDM, 01 PC 3He 1" Type: LND252/USA. **Keywords :** PC 3He, sensitivity, pulse height distribution spectra, gas amplification

Conference Title : ICBB 2015 : International Conference on Bioinformatics and Biomedicine

Conference Location : Istanbul, Türkiye

Conference Dates : May 21-22, 2015