Comparison between Transient Elastography (FibroScan) and Liver Biopsy for Diagnosis of Hepatic Fibrosis in Chronic Hepatitis C Genotype 4

Authors : Gamal Shiha, Seham Seif, Shahera Etreby, Khaled Zalata, Waleed Samir

Abstract : Background: Transient Elastography (TE; FibroScan®) is a non-invasive technique to assess liver fibrosis. Aim: To compare TE and liver biopsy in hepatitis C virus (HCV) patients, genotype IV and evaluate the effect of steatosis and schistosomiasis on FibroScan. Methods: The fibrosis stage (METAVIR Score) TE, was assessed in 519 patients. The diagnostic performance of FibroScan is assessed by calculating the area under the receiver operating characteristic curves (AUROCs). Results: The cut-off value of \geq F2 was 8.55 kPa, \geq F3 was 10.2 kPa and cirrhosis = F4 was 16.3 kPa. The positive predictive value and negative predictive value were 70.1% and 81.7% for the diagnosis of \geq F2, 62.6% and 96.22% for F \geq 3, and 27.7% and 100% for F4. No significant difference between schistosomiasis, steatosis degree and FibroScan measurements. Conclusion: Fibroscan could accurately predict liver fibrosis.

Keywords : chronic hepatitis C, FibroScan, liver biopsy, liver fibrosis

Conference Title : ICBB 2014 : International Conference on Bioinformatics and Biomedicine

Conference Location : Istanbul, Türkiye

Conference Dates : May 22-23, 2014