

Bi-Lateral Comparison between NIS-Egypt and NMISA-South Africa for the Calibration of an Optical Time Domain Reflectometer

Authors : Osama Terra, Mariesa Nel, Hatem Hussein

Abstract : Calibration of Optical Time Domain Reflectometer (OTDR) has a crucial role for the accurate determination of fault locations and the accurate calculation of loss budget of long-haul optical fibre links during installation and repair. A comparison has been made between the Egyptian National Institute for Standards (NIS-Egypt) and the National Metrology institute of South Africa (NMISA-South Africa) for the calibration of an OTDR. The distance and the attenuation scales of a transfer OTDR have been calibrated by both institutes using their standards according to the standard IEC 61746-1 (2009). The results of this comparison have been compiled in this report.

Keywords : OTDR calibration, recirculating loop, concatenated method, standard fiber

Conference Title : ICOMA 2016 : International Conference on Optical Metrology and Applications

Conference Location : Sydney, Australia

Conference Dates : December 15-16, 2016