## Disposable PANI-CeO2 Sensor for the Electrocatalytic Simultaneous Quantification of Amlodipine and Nebivolol

Authors : Nimisha Jadon, Rajeev Jain, Swati Sharma

**Abstract :** A chemically modified carbon paste sensor has been developed for the simultaneous determination of amlodipine (AML) and nebivolol (NBV). Carbon paste electrode (CPE) was fabricated by the addition of Gr/PANI-CeO2. Gr/PANI-CeO2/CPE has achieved excellent electrocatalytic activity and sensitivity. AML and NBV exhibited oxidation peaks at 0.70 and 0.90 V respectively on Gr/PANI-CeO2/CPE. The linearity range of AML and NBV was 0.1 to 1.6 µgmL-1 in BR buffer (pH 8.0). The Limit of detection (LOD) was 20.0 ngmL-1 for AML and 30.0 ngmL-1 for NBV and limit of quantification (LOQ) was 80.0 ngmL-1 for AML and 100 ngmL-1 for NBV respectively. These analyses were also determined in pharmaceutical formulation and human serum and good recovery was obtained for the developed method.

Keywords : amlodipine, nebivolol, square wave voltammetry, carbon paste electrode, simultaneous quantification

**Conference Title :** ICAEEE 2018 : International Conference on Advances in Electrochemistry and Electrochemical Engineering

**Conference Location :** Singapore, Singapore **Conference Dates :** March 22-23, 2018

1