

Determination of Vitamin C (Ascorbic Acid) in Orange Juices Product

Authors : Wanida Wonsawat

Abstract : This research describes a voltammetric approach to determine amounts of vitamin C (Ascorbic acid) in orange juice sample, using three screen printed electrode. The anodic currents of vitamin C were proportional to vitamin C concentration in the range of 0 – 10.0 mM with the limit of detection of 1.36 mM. The method was successfully employed with 2 μ L of the working solution dropped on the electrode surface. The proposed method was applied for the analysis of vitamin C in packed orange juice without sample purification or complexon of sample preparation step.

Keywords : ascorbic acid, vitamin C, juice, voltammetry

Conference Title : ICSDE 2014 : International Conference on Surface Design and Engineering

Conference Location : Paris, France

Conference Dates : June 26-27, 2014